



TANBead® Nucleic Acid Extraction Kit

OptiPure Blood DNA Auto Plate

(For use with the Maelstrom 9600 series)



W61EA46

(For Professional Use Only) V5

1. Intended Purpose

The TANBead® Nucleic Acid Extraction Kit is a nucleic acid purification kit based on magnetic bead technology by using with corresponding TANBead® Nucleic Acid Extractor, which can automatically isolate and purify DNA from human whole blood, frozen blood, and buffy coat. The purified DNA can be used with any downstream application employing PCR-based qualitative, semi-quantitative and quantitative assays. The kit is intended for use by technicians, physicians, and biologists with well-trained in molecular biological techniques, the techniques of magnetic bead purification and *in vitro* diagnostic procedures. Any diagnostic results generated by using the sample preparation procedure in conjunction with any downstream diagnostic assay should be interpreted related to other clinical or laboratory findings. The kit is not limited to any specific disorder, condition, or other additional accompanying diagnostics. It is applicable for all population.

2. The basic principle

The silicon dioxide layer coated on the magnetic beads can adsorb the negatively charged molecules to purify nucleic acids from samples.

3. Specification

Starting Materials	300 µL whole blood, frozen blood or buffy coat
Elution Volume	90~130 µL
Typical DNA yield	≥2 µg
Typical A260 / A280	≥1.7

4. Component Supplied with the Kit

Auto Plate	6	Auto Plate with reagent buffers
Proteinase K	1.0 mL x 1	Proteinase K
Elution Buffer	1.5 mL x 1	Nuclease-Free Water
Spin tips	96 tips	Spin tip assembled box
Protocol	1	Instruction guide for user

5. Auto Plate Content and Plate Position

Plate position	Buffer	Volume (µL)
1	Lysis Buffer	500
2	Washing Buffer 1	800
3	Magnetic Beads	800
4	Washing Buffer 2	800
5	Washing Buffer 2	800
6	Elution Buffer	130
7	N / A	N / A
8	Spin tip	-

6. Kit Storage and Shelf Life

- Components under room temperature (15~35°C) can be stored until the expiration date labeled on the box.
- The proteinase K is transported at room temperature. When received, please store proteinase K at 2~8°C.

7. Precautions

- It can be used for *in vitro* diagnostic use.
- When the temperature is below 20°C, place the reagent plate in an oven (preheated 42~60°C) from 5 to 10 minutes.
- Avoid vigorous shaking, in order to avoid excessive formation of foam.
- Carefully remove aluminum foil to avoid splashing.
- Do not expose the opened reagents or Auto Plates / Auto Tubes to air. The evaporation would lead to pH change, or effect on the extraction effectiveness.
- The reagents are all colorless and transparent. Colored reagents indicate contamination, please replace it with a fresh Auto Plates / Auto Tubes before proceeding.

- Please check the integrity of the Auto Plates / Auto Tubes and remember to mount the spin tips into the appropriate position of the suitable instrument before operating them.
- Please wear a mask and disposable gloves when handling.
- Use sterile consumables to avoid nuclease contamination.
- Reagent solution contains guanidine salt, avoid using bleach containing detergent.
- Avoid eyes, skin, and clothing contact with reagents. In case of any contact, flush with flowing water.
- If any serious incident occurs, please report to the manufacturer and the competent authority of the member state in which the user and / or the patient is established.

8. Materials required, Not Supplied

- TANBead® Nucleic Acid Extraction System
Model: Maelstrom 9600 series (non-sterile)
- Disposable gloves
- Scissors, utility knives
- Micropipette, disposable tips (10 µL / 200 µL / 1000 µL)
- 1.5 mL microcentrifuge tube
- 15 mL / 50 mL conical tube

9. Sample Collection, Transportation, and Storage

■ Sample collection and storage

- Whole blood and buffy coat collection
 - Whole blood and buffy coat specimens must be obtained from sodium citrate or EDTA collection tubes.
- Specimen storage
 - Fresh whole blood specimens can be stored at room temperature for 6 hours.
 - After centrifugation, the buffy coat sample can be stored at
 - 2~8°C up to 7 days.
 - 20°C for long-term preservation.

■ Specimen transportation

Transportation of whole blood, buffy coat specimen should follow specific blood transportation related law. Whole blood sample and buffy coat should be kept between 2~25°C during transportation and separate buffy coat within 6 hours.

10. Nucleic Acids Extraction Protocol

- Carefully remove the aluminum foil on the Auto Plates.
- Add **300 µL blood** into wells of **plate #1**.
- Add **10 µL Proteinase K** into wells of **plate #1**.
- Select the program "**61E**". The parameters are given in the following section.
- Follow the guide shown on the screen and place plates carefully. Make sure that the chamfer of the plate faces toward the same direction.
- Carefully remove the Auto Plates when the program is finished.
- Use micropipette to transfer the purified nucleic acids from **plate #6** to a clean tube.
- Discard used Auto Plates and spin tips into the waste recycling bin.

11. Program

■ Maelstrom 9600 series

Program Name: 61E								
Plate	1	2	3	4	5	6	7	8
Volume(μL)	800	800	800	800	800	100	-	-
Keep Temp.	60	40	40	0	0	30	-	-
Action	For.	For.	For.	For.	For.	For.	-	-
Name	LB	WB1	MB	WB2	WB2	EB	-	TIP
Step	Plate	Temp. (°C)	Mixing (min)	Mixing (rpm)	Collect (sec)	Vapor (min)	Pause	
1	3	45	0.5	3000	30	0.0	Off	
2	2	45	0.5	3000	30	0.0	Off	
3	1	65	15.0	3000	30	0.0	Off	
4	2	60	5.0	3000	30	0.0	Off	
5	3	45	3.0	3000	30	0.0	Off	
6	4	0	3.0	3000	30	0.0	Off	
7	5	0	3.0	3000	30	5.0	Off	
8	6	40	5.0	3000	60	0.0	Off	
9	5	0	0.2	3000	0	0.0	Off	

⚠ Temperature set as "0" represents room temperature!

■ Maelstrom 9610 series

Program Name: 61E								
Plate	1	2	3	4	5	6	7	8
Volume(μL)	800	800	800	800	800	100	-	-
Keep Temp.	60	40	40	-	-	30	-	-
Action	For.	For.	For.	For.	For.	For.	-	-
Name	LB	WB1	MB	WB2	WB2	EB	-	TIP
Step	Plate	Temp. (°C)	Mixing (min)	Mixing (rpm)	Collect (sec)	Vapor (min)	Pause	
1	3	45	0.5	3000	30	0.0	Off	
2	2	45	0.5	3000	30	0.0	Off	
3	1	65	15.0	3000	30	0.0	Off	
4	2	60	5.0	3000	30	0.0	Off	
5	3	45	3.0	3000	30	0.0	Off	
6	4	-	3.0	3000	30	0.0	Off	
7	5	-	3.0	3000	30	5.0	Off	
8	6	40	5.0	3000	60	0.0	Off	
9	5	-	0.2	3000	0	0.0	Off	

⚠ Temperature set as "25" represents room temperature!

12. Reagent performance

■ Repeatability

Under repeatability conditions where nucleic acids are extracted with the same reagent kit on the same source samples by the same operator. The coefficient of variation of nucleic acid extraction concentration is less than 5%.















■ Reproducibility

A five-day reproducibility test was carried out with the same source samples for 5 consecutive days with the same reagent kit by different operators. The coefficient of variation of nucleic acid extraction concentration is less than 5%.

■ The stability of extracted DNA

Storage Conditions	DNA stability
-80°C	Over 90 days
-20°C	28 days
4°C	14 days
25°C	2 days
Freeze-thaw	5 times

13. Explanation of Symbols

	Manufacturer		Consult instructions for use
	Temperature limit		Contains sufficient for test
	CE mark		In vitro diagnostic medical use
	Catalogue number		Caution
	Batch code		Non-sterile
	Do not re-use		Keep away from sunlight
	Date of manufacture		Use-by date

EC REP

mdi Europa GmbH, Langenhagener Str. 71, 30855 Langenhagen, Germany

14. Post-market surveillance conclusion

After a risk assessment and clinical evaluation assessment, when weighing the benefits of medical device, patients, and the risks associated with the use of the device, the risk is acceptable. The post-market surveillance report shows that no death or serious adverse events occurred.